

DECLARATION OF CONFORMITY

Replaces version dated: -

Valid until the issue of next version
of this document.

We,

OneMed Group Oy,
Metsäläntie 20,
FI-00320 Helsinki, Finland,

SRN FI-MF-000000642,

Domicile Helsinki,
Business ID 2039640-1,

declare under our sole responsibility that following products are in conformity with Regulation (EU) 2017/745 of the European Parliament and of the Council of 5 April 2017 on medical devices as well as Regulation (EU) 2016/425 of the European Parliament and of the Council of 9 March 2016 on personal protective equipment.

Classification:

Class I according to Annex VIII of the **Regulation (EU) 2017/745 on medical devices,**

Category III according to the **Regulation (EU) 2016/425 on personal protective equipment**

Intended Purpose (As per Regulation (EU) 2017/745 on medical devices):

Nitrile examination gloves are for medical purposes intended to be worn on care giver's hands to prevent cross-contamination between care giver and a patient. The gloves are non-sterile. The gloves are for single use. The device may be used by health care professional or lay person.

List of devices under Basic UDI-DI 6438129B0001GU:

Item number (REF)	Product name
6530	evercare [®] Nitrile Medical Examination Gloves, Accelerator-free, blue, Neptune, size XS
6531	evercare [®] Nitrile Medical Examination Gloves, Accelerator-free, blue, Neptune, size S
6532	evercare [®] Nitrile Medical Examination Gloves, Accelerator-free, blue, Neptune, size M
6533	evercare [®] Nitrile Medical Examination Gloves, Accelerator-free, blue, Neptune, size L
6534	evercare [®] Nitrile Medical Examination Gloves, Accelerator-free, blue, Neptune, size XL

These items are in conformity with following European standards and common specifications¹:

EN ISO 13485	Medical Devices – Quality management systems – Requirements for regulatory purposes
EN ISO 14971	Medical Devices – Application of risk management to medical devices
EN ISO 20417	Information supplied by the manufacturer with medical devices
EN ISO 15223-1	Medical Devices – Symbols to be used with medical device labels, labelling and information to be supplied. Part 1: General requirements
EN 420	Protective glove. General requirements and test method
EN 455-1	Medical gloves for single use. Part 1 – Requirements and testing for freedom from holes
EN 455-2	Medical gloves for single use. Part 2 – Requirements and testing for physical properties
EN 455-3	Medical gloves for single use. Part 3 – Requirements and testing for biological evaluation
EN 455-4	Medical gloves for single use. Part 4 – Requirements and testing for shelf life determination
EN ISO 374-1	Protective gloves against chemicals and micro-organism. Part 1 - Terminology and performance requirements for chemical risks

EN ISO 374-2	Protective gloves against dangerous chemicals and micro-organism – Part 2: Determination of resistance to penetration
EN 374-4	Protective gloves against dangerous chemicals and micro-organism – Part 4: Determination of resistance to degradation by chemicals
EN ISO 374-5	Protective gloves against dangerous chemicals and micro-organisms. Part 5 - Terminology and performance requirements for micro-organisms risks

¹Relevant standards, latest applied revisions and common specifications are listed in *GEN-079 Review of regulations and standards*

The notified body 2777, SATRA Technology Europe Limited Bracetown, Business Park, Clonee, D15YN2P, Republic of Ireland performed the EU type-examination (Module B) and issued the EU type-examination certificate 2777/13883-03/E02-01. The product(s) is (are) subject to the conformity assessment procedure Module D under surveillance of the same notified body. Type C glove according to EN ISO 374-1:2016+A1:2018

Place and date of issue

Helsinki, 30.01.2026

Name and signature of the authorized person


Hilleriikka Vaho
Regulatory Affairs Manager
OneMed Group Oy